



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/525,044

08/01/2005

Giorgio Mari

P70417US0

5665

136 7590 12/17/2008

JACOBSON HOLMAN PLLC
400 SEVENTH STREET N.W.
SUITE 600
WASHINGTON, DC 20004

EXAMINER

CHRISTIAN, MARJORIE ELLEN

ART UNIT

PAPER NUMBER

1797

MAIL DATE

DELIVERY MODE

12/17/2008

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary	Application No. 10/525,044	Applicant(s) MARI ET AL.	
	Examiner MARJORIE CHRISTIAN	Art Unit 1797	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 18 February 2005.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-18 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-18 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☒ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☒ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____ |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date <u>10/17/2005</u> . | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Summary

1. This is the initial Office action based on the application filed February 18th, 2005.
2. The petition filed 3/6/2006 has been entered.
3. Claims 11-18 are pending and have been fully considered.

Priority

4. Receipt is acknowledged of papers submitted under 35 U.S.C. 119(a)-(d), which papers have been placed of record in the file.

Claim Rejections - 35 USC § 112


5. Regarding claim 14, the phrase "e.g." renders the claim indefinite because it is unclear whether the limitation following the phrase is part of the claimed invention. See MPEP § 2173.05(d).

Claim Rejections - 35 USC § 102

6. **Claims 1-3, 7-13, 16-18 are rejected under 35 U.S.C. 102(b) as being anticipated by WO00/54873, BORMANN et al. (hereinafter BORMANN).**

As to Claims 1-2, BORMANN discloses a filter device for the depletion of the leukocyte content from blood products (Abstract) comprising a housing (Fig. 1, Ref. 100) with an inlet (20) and an outlet (30) port and, within said housing, two porous elements (1, 2) *adapted for* removing leukocytes, each porous element comprising one

or more layers of filtering material (Pg. 11, Lines 13-14), wherein said at least two porous elements have a different hydrophilicity (Pg. 12, Lines 5-6), the first porous element has a higher hydrophilicity than the successive filter element in the direction of flow, from inlet to outlet in the direction of blood flow (Page 12, Lines 6-9, Figs. 1-3).



As to Claim 3, BORMANN discloses each porous element comprises at least two adjacent layers of filtering material (Pg. 11, Lines 12-14).

As to Claim 7, BORMANN discloses said porous elements are made of fibers consisting of polyolefin (Pg. 11, Line 3) coated with a hydrophilic polymer (Pg. 11, Lines 16-22).

As to Claim 8, BORMANN discloses said hydrophilic polymer is acrylic polymers (Pg. 14, Lines 13-19).

As to Claim 9, BORMANN discloses a filter device including porous elements made of polybutylene terephthalate (Pg. 11, Lines 1-5), where first porous element can be coated with a hydrophilic polymer and the second element can be uncoated (Pg. 11, Lines 15-28).

As to Claim 10, BORMANN discloses the porous elements are arranged in the filter device according to a decreasing value of the CWST (Pg. 12, Lines 2-5), from inlet to outlet (Pg. 12, Lines 10-15, Figs. 1-3).

As to Claims 11-12, BORMANN discloses the difference between the hydrophilicity of the inlet (1) and outlet porous element (2), as measured by the CWST of the material is from 10dyn/cm to 20dyn/cm (Pg. 12, Lines 10-15).

As to Claim 13, BORMANN discloses the first porous element has hydrophilicity, as measured by the CWST, of greater than 63dyn/cm (Pg. 12, Lines 6-9).

As to Claim 16, BORMANN discloses a blood bag device for the separation of blood into leukocyte depleted blood components (Fig. 4) comprising a first bag (51), in fluid flow communication with a second bag (50) through a leukocyte filter device (100) according to claim 1 (see 102(b) rejection of Claim 1).

As to Claim 17, BORMANN discloses a method for leukocyte depletion of blood products comprising feeding said blood product through a filter device (BORMANN, Claim 29) according to claim 1 (see 102(b) rejection of Claim 1).

As to Claim 18, BORMANN discloses said blood product is plasma (BORMANN, Claim 9).

Claim Rejections - 35 USC § 103

7. **Claims 4-5 are rejected under 35 U.S.C. 103(a) as obvious over WO00/54873, BORMANN et al. (hereinafter BORMANN) in view of US Patent No. 4,925,572, PALL (hereinafter PALL).**

As to Claim 4, BORMANN discloses layers of filtering material in the porous element (Pg. 11, Lines 12-14). BORMANN does not appear to explicitly disclose that the layers are made of the same material and subsequently have the same hydrophilicity properties. However, PALL discloses the use of multiple layers in a filter element made of the same material (C28/L7-14).

At the time of the invention, it would have been obvious to one of ordinary skill in the art to modify the filter material of the layers of BORMANN to include layers of the same material of PALL. The motivation would have been to efficiently remove micro-aggregates and partially free the fluid from leukocytes (C15/L60-63). Therefore, the invention as a whole would have been *prima facie* obvious to one of ordinary skill in the art at the time the invention was made.

As to Claim 5, BORMANN discloses that the filter elements can have different and varied pore structures (Pg. 16, Lines 4-5). Specifically, PALL discloses layers having a decreasing pore size from inlet to outlet (C15/L55-59).

8. **Claim 6 is rejected under 35 USC 103 (a) as being obvious over WO00/54873, BORMANN et al. (hereinafter BORMANN) in view of US Patent No. 5,298,165, OKA et al. (hereinafter OKA).**

As to Claim 6, BORMANN discloses that the filter elements can have different and varied pore structures (Pg. 16, Lines 4-5). BORMANN does not appear to explicitly disclose that the filter elements have decreasing pore size from inlet to outlet. However, OKA discloses two layers have a decreasing pore size from inlet to outlet, where the

Art Unit: 1797

filtering material has a pore size higher than its successive porous element (OKA, Claims 1-2).

At the time of the invention, it would have been obvious to one of ordinary skill in the art to modify the different pore structures of the filter elements of BORMANN to include the decreasing pore size from inlet to outlet of the porous elements of OKA. The motivation would have been to more effectively remove leukocytes from a blood product (Abstract). Therefore, the invention as a whole would have been *prima facie* obvious to one of ordinary skill in the art at the time the invention was made.


9. **Claims 14-15 are rejected under 35 USC 103 (a) as being obvious over WO00/54873, BORMANN et al. (hereinafter BORMANN) in view of US Patent No. 5,190,657, HEAGLE et al. (hereinafter HEAGLE).**

As to Claim 14, BORMANN discloses the filter device and filtration elements for leukocyte depletion as shown in the 102(b) rejection of Claim 1. BORMANN does not appear to explicitly disclose the use of a filtration element which is not adapted for leukocyte removal. However, HEAGLE discloses a pre-filter (Fig. 6, Ref. 15) which is not adapted for leukocyte removal (C14/L9-16).

At the time of the invention, it would have been obvious to one of ordinary skill in the art to modify the filtration device for leukocyte depletion of BORMANN to include the pre-filter element which is not adapted for leukocyte removal of HEAGLE. The suggestion would have been to remove large agglomerates prior to contact with the

Art Unit: 1797

filter element (C14/L9-16). Therefore, the invention as a whole would have been *prima facie* obvious to one of ordinary skill in the art at the time the invention was made.



As to Claim 15, HEAGLE discloses said filter elements not adapted for leukocyte removal (15) are located closer to the inlet (18) than said elements adapted for leukocyte removal (1).

Conclusion

10. The prior art made of record and not relied upon is considered pertinent to applicant's disclosure.

- US Patent No. 4,880,548 as it discloses a CWST of the filter element of 90dyn/cm;
- US Patent No. 4,985,153 as it discloses layers of filter material for leukocyte depletion; and
- US Patent No. 5,229,012 discloses as it discloses a CWST between 10-20dyn/cm for a leukocyte depletion filter.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to MARJORIE CHRISTIAN whose telephone number is (571)270-5544. The examiner can normally be reached on Monday through Thursday 7-5pm (Fridays off).

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, David R. Sample can be reached on (571)272-1376. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

MC

/Krishnan S Menon/
Primary Examiner, Art Unit 1797